

35/09.14/6.2

Н

Technica	al Data Sheet	Rosenbo		
RPC-2.92	Mismatch Jack	02K150-060S3		

Electrical data	
Frequency range VSWR	DC to 40 GHz 1.2 ± 0.06, DC to 4 GHz 1.2 ± 0.08, 4 GHz to 18 GHz
DC Resistance Power handling	1.2 ± 0.10 , 18 GHz to 40 GHz 60 Ω ≤ 0.5 W
Mechanical data Mating cycles Maximum torque Recommended torque Gauge	≥ 500 1.70 Nm 0.90 Nm 0.00 mm to 0.03 mm
Environmental data Operating temperature range ¹ Rated temperature range of use ² Storage temperature range	+20 °C to +26 °C 0 °C to +50 °C -40 °C to +85 °C
RoHS	compliant
 ¹ Temperature range over which these specification are ² This range is underneath and above the operating ten and could be used without damage. 	

Declaration of calibration options

Factory Calibration

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, traceable to national / international standards.

Accredited Calibration

Optional this calibration standard can be delivered with an Accredited Calibration (DAkkS) having the highest confidence in the traceability. The DAkkS Calibration Certificate issued reports individual calibration results in a complex format, traceable to national / international standards. Calibration results are reported in a dense data set file. The uncertainties are smaller than in a Factory Calibration.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation

12 months

Packing

Standard Weight 1 pce in box 8.0 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date		Rev.	Engineering change number	Name	Date
Herbert Babinger	09.12.14	Lars Ramtke	08.10.19		g00	19-1910	Marion Striegler	08.10.19
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de					Tel.	Tel. : +49 8684 18-0		Page
					Email : info@rosenberger.de		2/2	